

THE FOLLOWING INFORMATION IS LEGALLY RELEVANT AND IS EQUIVALENT TO THE WELMEC APPROACH:

In Estonia, all calculations for weighing instruments, where applicable, have to be carried out for value  $g = 9,817823 \text{ m/s}^2$ .

Title: Regulation of Minister of Economic Affairs and Communications No 107 from 21.04.2004 on non-automatic weighing instruments.