

WELMEC

European Cooperation in Legal Metrology

Measuring Instruments Directive 2014/32/EU
Continuous Totalisers
Corresponding Tables
OIML R 50-1/-2 : 2014 – Annex VIII, Chapter V



WELMEC

European Cooperation in Legal Metrology

WELMEC is a cooperation between the legal metrology authorities of the Member States of the European Union and EFTA.

This document is one of a number of Guides published by WELMEC to provide guidance to manufacturers of measuring instruments and to Notified Bodies responsible for conformity assessment of their products.

The Guides are purely advisory and do not themselves impose any restrictions or additional technical requirements beyond those contained in relevant EU Directives.

Alternative approaches may be acceptable, but the guidance provided in this document represents the considered view of WELMEC as to the best practice to be followed.

Published by:
WELMEC Secretariat

E-mail: secretary@welmec.org
Website: www.welmec.org

Continuous Totalisers
Corresponding Table 2014/32/EU vs. OIML R 50-1:2014 and R50-2:2014 (E)

Notes

1. The column “Comments” indicates, if necessary, the relevant text of OIML R 50-1 and R 50-2 related explanations concerning the compliance with the relevant requirements of Directive 2014/32/EU.
2. The column “Conclusion” gives the conclusion on the compliance between OIML R 50-1 and R 50-2 and the relevant requirements of Directive 2014/32/EU.

The indication “Covered” means that:

- the requirement of OIML R 50-1 and R 50-2 is identical to the one of Directive 2014/32/EU, or
- the requirement of OIML R 50-1 and R 50-2 is more severe than the one of Directive 2014/32/EU, or all the requirements of OIML R 50-1 fulfil requirements of Directive 2014/32/EU (even when Directive 2014/32/EU allows other alternatives),
- The indication “Partially covered” means that the requirement is not fully covered or R 50 allows possibilities which are not foreseen in the directive.

The indication “Not Covered” means that the requirement in Directive 2014/32/EU is either not compatible with the relevant OIML R 50-1 and R 50-2 requirement or not included in OIML R 50-1 and R 50-2.

The indication “Not Relevant” means that the requirement in Annex I of Directive 2014/32/EU is not relevant for continuous totalisers.

The text in *italic* is an extract from the relevant clause of the OIML Recommendation.

Directive 2014/32/EU Essential requirements of Annex I and Annex VIII Automatic Weighing Instruments (MI-006)	OIML R50-1/-2 : 2014(E)	Comments	Conclusions
ANNEX 1			
1. Allowable Errors			
1.1. Under rated operating conditions and in the absence of a disturbance, the error of measurement shall not exceed the maximum permissible error (MPE) value as laid down in the appropriate instrument-specific requirements. Unless stated otherwise in the instrument-specific annexes, MPE is expressed as a bilateral value of the deviation from the true measurement value	R50-1, 4.2 R50-1, 3.2.1 & 3.2.2		Covered
1.2. Under rated operating conditions and in the presence of a disturbance, the performance requirement shall be as laid down in the appropriate instrument specific requirements. Where the instrument is intended to be used in a specified permanent continuous electromagnetic field the permitted performance during the radiated electromagnetic field-amplitude modulated test shall be within MPE.	R50-1, 5.1.1, 5.3, 5.5.2	The frequency range up to 2 GHz is not sufficient according to D11 (3 GHz).	Partially covered
1.3. The manufacturer shall specify the climatic, mechanical and electromagnetic environments in which the instrument is intended to be used, power supply and other influence quantities likely to affect its accuracy, taking account of the requirements laid down in the appropriate instrument-specific annexes.	R50-1, 6.1.1		Covered
1.3.1. <i>Climatic environments</i> The manufacturer shall specify the upper temperature limit and the lower temperature limit from any of the values in Table 1 unless otherwise specified in the annexes III to XII, and indicate whether the instrument is designed for condensing or non-condensing humidity as well as the intended location for the instrument, i.e. open or closed.	R50-1, 3.7.4.1 R50-1, 5.5.1 R50-1, 6.1.1	No reference to location Covered except for-intended location	Partially covered

Directive 2014/32/EU Essential requirements of Annex I and Annex VIII Automatic Weighing Instruments (MI-006)					OIML R50-1/-2 : 2014(E)	Comments	Conclusions
Table 1					R50-1, 3.7.4.1	-10 to +40 °C (others possible) Minimum range of 30 °C	Covered
	Temperature Limits						
Upper temperature limit	30 °C	40 °C	55 °C	70 °C			
Lower temperature limit	5 °C	- 10 °C	- 25 °C	- 40 °C			
1.3.2. (a) Mechanical environments are classified into classes M1 to M3 as described below. M1 This class applies to instruments used in locations with vibration and shocks of low significance, e.g. for instruments fastened to light supporting structures subject to negligible vibrations and shocks transmitted from local blasting or pile-driving activities, slamming doors, etc. M2 This class applies to instruments used in locations with significant or high levels of vibration and shock, e.g. transmitted from machines and passing vehicles in the vicinity or adjacent to heavy machines, conveyor belts, etc. M3 This class applies to instruments used in locations where the level of vibration and shock is high and very high, e.g. for instruments mounted directly on machines, conveyor belts, etc. (b) The following influence quantities shall be considered in relation with mechanical environments: – Vibration; – Mechanical shock.						Classes not covered in R50-1:2014	Not relevant
						Not relevant according to annex VIII Chapter 1 §1.3	Not relevant
						Not relevant according to annex VIII Chapter 1 §1.3	Not relevant
						Not relevant according to annex VIII Chapter 1 §1.3	Not relevant
						Not relevant according to annex VIII Chapter 1 §1.3	Not relevant

Directive 2014/32/EU Essential requirements of Annex I and Annex VIII Automatic Weighing Instruments (MI-006)	OIML R50-1/-2 : 2014(E)	Comments	Conclusions
1.3.3. (a) Electromagnetic environments are classified into classes E1, E2 or E3 as described below, unless otherwise laid down in the appropriate instrument-specific annexes.		Definitions of E1, E2 & E3 are not mentioned in the recommendation	Not covered
E1 This class applies to instruments used in locations with electromagnetic disturbances corresponding to those likely to be found in residential, commercial and light industrial buildings.		Definitions of E1, E2 & E3 are not mentioned in the recommendation	Not covered
E2 This class applies to instruments used in locations with electromagnetic disturbances corresponding to those likely to be found in other industrial buildings.		Definitions of E1, E2 & E3 are not mentioned in the recommendation	Not covered
E3 This class applies to instruments supplied by the battery of a vehicle. Such instruments shall comply with the requirements of E2 and the following additional requirements:		Definitions of E1, E2 & E3 are not mentioned in the recommendation	Not covered
– voltage reductions caused by energising the starter-motor circuits of internal combustion engines		Voltage reduction test not mentioned. Not covered by R50-1 and R50-2 but it is in D11:2013 OIML D11:2013 (14.2, Table 40)	Not covered
– load dump transients occurring in the event of a discharged battery being disconnected while the engine is running.		Load dump is not mentioned. Not covered by R50-1 and R50-2 but it is in D11:2013 OIML D11:2013 (14.2, Table 41)	Not covered
(b) The following influence quantities shall be considered in relation with electromagnetic environments:	R50-1, 5.1.1, 5.5.2		Partially covered

Directive 2014/32/EU Essential requirements of Annex I and Annex VIII Automatic Weighing Instruments (MI-006)	OIML R50-1/-2 : 2014(E)	Comments	Conclusions
– Voltage interruptions,	R50-2 7.3.1	Covered on the provision that the relevant severity level specified in OIML D11:2013 is used OIML D11:2013 (12.3, Table 23). For E1 use severity level 1. For E2 use severity level 2.	Partially covered
– Short voltage reductions,	R50-2 7.3.1	Covered on the provision that the relevant severity level specified in OIML D11:2013 is used OIML D11:2013 (12.3, Table 23). For E1 use severity level 1. For E2 use severity level 2.	Partially covered
– Voltage transients on supply lines and/or signal lines,	R50-2 7.3.2 R50-2 7.3.24	Covered on the provision that the relevant severity level specified in OIML D11:2013 is used OIML D11:2013 (12.3, Table 26). For E1 use severity level 2. For E2 use severity level 3. OIML D11:2013 (12.4, Table 28). For E1 use severity level 2. For E2 use severity level 3.	Partially covered
– Electrostatic discharges,	R50-2 7.3.4		Covered
– Radio frequency electromagnetic fields,	R50-2 7.3.5.1	Covered on the provision that the relevant severity level specified in OIML D11:2013 is used OIML D11:2013 (13.2, Tables 32, 33 and 34*). For E1 use severity level 2. For E2 use severity level 3. * limit frequency range in Table 34 to 3 GHz.	Partially covered

Directive 2014/32/EU Essential requirements of Annex I and Annex VIII Automatic Weighing Instruments (MI-006)	OIML R50-1/-2 : 2014(E)	Comments	Conclusions
<ul style="list-style-type: none"> Conducted radio frequency electromagnetic fields on supply lines and/or signal lines, 	R50-2 7.3.5.2	Covered on the provision that the relevant severity level specified in OIML D11:2013 is used OIML D11:2013 (13.2, Table 31). For E1 use severity level 2. For E2 use severity level 3.	Partially covered
<ul style="list-style-type: none"> Surges on supply lines and/or signal lines. 	R50-2 7.3.3 R50-2 7.3.3	Covered on the provision that the relevant severity level specified in OIML D11:2013 is used OIML D11:2013 (12.3, Table 27). For E1 use severity level 2. For E2 use severity level 3. OIML D11:2013 (12.4, Table 29). For E1 use severity level 2. For E2 use severity level 3.	Partially covered
1.3.4. Other influence quantities to be considered, where appropriate, are:		Not all quantities are covered	Partially covered
<ul style="list-style-type: none"> Voltage variation, 	R50-1, 3.7.4.3 & 5.5.4 R50-2, 7.2.4 & 7.2.5 R50-2, 6.5 & 6.6	Only covered for E1 and E2	Partially Covered
<ul style="list-style-type: none"> Mains frequency variation, 		Not covered by R50-1 and R50-2 but in D11:2013 OIML D11:2013 (12.2, Table 21).	Not covered
<ul style="list-style-type: none"> Power frequency magnetic fields, 		Not covered by R50-1 and R50-2 but in D11:2013 OIML D11:2013 (13.1, Table 30).	Not covered

Directive 2014/32/EU Essential requirements of Annex I and Annex VIII Automatic Weighing Instruments (MI-006)	OIML R50-1/-2 : 2014(E)	Comments	Conclusions
<p>– Any other quantity likely to influence in a significant way the accuracy of the instrument.</p>	<p>R50-1, 3.7.5.4 R50-2, 8.4</p> <p>R50-1, 5.5.3 R50-2, 3.5 & 5.2</p> <p>R50-1, 3.7.2 R50-2, 5.4.2</p> <p>R50-1, 3.7.1 R50-2, 5.4.1</p> <p>R50-1, 3.7.4.2 R50-2, 7.2.2</p>	<p>Short-term / long term stability of zero</p> <p>Warm-up time</p> <p>Eccentric loading</p> <p>Variation of simulation speed</p> <p>Temperature effect at zero flowrate</p>	<p>Covered</p>
<p>1.4. When carrying out the tests as envisaged in this Directive, the following points shall apply:</p>			
<p>1.4.1. <i>Basic rules for testing and the determination of errors</i></p> <p>Essential requirements specified in 1.1 and 1.2 shall be verified for each relevant influence quantity. Unless otherwise specified in the appropriate instrument-specific annex, these essential requirements apply when each influence quantity is applied and its effect evaluated separately, all other influence quantities being kept relatively constant at their reference value.</p> <p>Metrological tests shall be carried out during or after the application of the influence quantity, whichever condition corresponds to the normal operational status of the instrument when that influence quantity is likely to occur.</p>	<p>R50-2, 7.1</p> <p>R50-2</p>	<p>Conditions are listed in each test</p>	<p>Covered</p> <p>Covered</p>

Directive 2014/32/EU Essential requirements of Annex I and Annex VIII Automatic Weighing Instruments (MI-006)	OIML R50-1/-2 : 2014(E)	Comments	Conclusions
<p><i>1.4.2. Ambient humidity</i></p> <p>(a) According to the climatic operating environment in which the instrument is intended to be used either the damp heat-steady state (non-condensing) or damp heat cyclic (condensing) test may be appropriate.</p> <p>(b) The damp heat cyclic test is appropriate where condensation is important or when penetration of vapour will be accelerated by the effect of breathing. In conditions where non-condensing humidity is a factor the damp-heat steady state is appropriate.</p>	<p>R50-1, 5.5.1 R50-2, 7.2.3</p>		<p>Covered</p>
<p>2. Reproducibility</p> <p>The application of the same measurand in a different location or by a different user, all other conditions being the same, shall result in the close agreement of successive measurements. The difference between the measurement results shall be small when compared with the MPE</p>	<p>R50</p>	<p>Covered provided all requirements and tests should be performed and in accordance with R 50.</p>	<p>Partially covered</p>
<p>3. Repeatability</p> <p>The application of the same measurand under the same conditions of measurement shall result in the close agreement of successive measurements. The difference between the measurement results shall be small when compared with the MPE.</p>	<p>R50-1, 3.7.5.1 & 3.8.1 R50-2, 8.1</p>		<p>Covered</p>
<p>4. Discrimination and Sensitivity</p> <p>A measuring instrument shall be sufficiently sensitive and the discrimination threshold shall be sufficiently low for the intended measurement task.</p>	<p>R50-1, 3.7.5.2, 3.7.5.3, 3.8.3 R50-2, 8.2, 8.3, 9.1.1</p>		<p>Covered</p>

Directive 2014/32/EU Essential requirements of Annex I and Annex VIII Automatic Weighing Instruments (MI-006)	OIML R50-1/-2 : 2014(E)	Comments	Conclusions
5. Durability A measuring instrument shall be designed to maintain an adequate stability of its metrological characteristics over a period of time estimated by the manufacturer, provided that it is properly installed, maintained and used according to the manufacturer's instruction when in the environmental conditions for which it is intended.	R50-1, 3.9		Covered
6. Reliability A measuring instrument shall be designed to reduce as far as possible the effect of a defect that would lead to an inaccurate measurement result, unless the presence of such a defect is obvious.	R50-1, 4.3.1 to 4.3.6 R50-1, 4.4.1 R50-1, 4.7, 4.8 R50-1, 5.1.1, 5.3		Covered
7. Suitability			
7.1. A measuring instrument shall have no feature likely to facilitate fraudulent use, whereas possibilities for unintentional misuse shall be minimal.	R50-1, 3.3 R50-1, 4.1, 4.3, 4.3.3 R50-1, 4.4.6 R50-1, 4.3.7		Covered
7.2. A measuring instrument shall be suitable for its intended use taking account of the practical working conditions and shall not require unreasonable demands of the user in order to obtain a correct measurement result.	R50-1, 4.1		Covered
7.3. The errors of a utility measuring instrument at flows or currents outside the controlled range shall not be unduly biased.			Not relevant
7.4. Where a measuring instrument is designed for the measurement of values of the measurand that are constant over time, the measuring instrument shall be insensitive to small fluctuations of the value of the measurand, or shall take appropriate action.			Not relevant

Directive 2014/32/EU Essential requirements of Annex I and Annex VIII Automatic Weighing Instruments (MI-006)	OIML R50-1/-2 : 2014(E)	Comments	Conclusions
7.5. A measuring instrument shall be robust and its materials of construction shall be suitable for the conditions in which it is intended to be used.	R50-1, 4.1, 4.8		Covered
7.6. A measuring instrument shall be designed so as to allow the control of the measuring tasks after the instrument has been placed on the market and put into use. If necessary, special equipment or software for this control shall be part of the instrument. The test procedure shall be described in the operation manual. When a measuring instrument has associated software which provides other functions besides the measuring function, the software that is critical for the metrological characteristics shall be identifiable and shall not be inadmissibly influenced by the associated software.	R50-1, 6.1.3.1, 6.2.2.1, 7.1 R50-2, 9 & 10 R50-1, 5.8 R50-1, 2.2.11.6 R50-1, 4.3.7.1		Covered Covered
8. Protection against corruption			
8.1. The metrological characteristics of a measuring instrument shall not be influenced in any inadmissible way by the connection to it of another device, by any feature of the connected device itself or by any remote device that communicates with the measuring instrument.	R50-1, 5.6		Covered
8.2. A hardware component that is critical for metrological characteristics shall be designed so that it can be secured. Security measures foreseen shall provide for evidence of an intervention.	R50-1, 4.3.7		Covered
8.3. Software that is critical for metrological characteristics shall be identified as such and shall be secured. Software identification shall be easily provided by the measuring instrument. Evidence of an intervention shall be available for a reasonable period of time.	R50-1, 5.8, 4.3.7		Covered

Directive 2014/32/EU Essential requirements of Annex I and Annex VIII Automatic Weighing Instruments (MI-006)	OIML R50-1/-2 : 2014(E)	Comments	Conclusions
8.4. Measurement data, software that is critical for measurement characteristics and metrologically important parameters stored or transmitted shall be adequately protected against accidental or intentional corruption.	R50-1, 4.3.7.1 R50-1, 5.8		Covered
8.5. For utility measuring instruments the display of the total quantity supplied or the displays from which the total quantity supplied can be derived, whole or partial reference to which is the basis for payment, shall not be able to be reset during use.			Not relevant
9. Information to be borne by and to accompany the instrument			
9.1. A measuring instrument shall bear the following inscriptions: (a) manufacturer's name, registered trade name or registered trade mark; (b) information in respect of its accuracy, and, where applicable: (c) information in respect of the conditions of use; (d) measuring capacity; (e) measuring range; (f) identity marking; (g) number of the EU-type examination certificate or the EU design examination certificate; (h) information whether or not additional devices providing metrological results comply with the provisions of this Directive on legal metrological control.	R50-1, 4.9		Covered
9.2. An instrument of dimensions too small or of too sensitive a composition to allow it to bear the relevant information shall have its packaging, if any, and the accompanying documents required by the provisions of this Directive suitably marked.			Not relevant

Directive 2014/32/EU Essential requirements of Annex I and Annex VIII Automatic Weighing Instruments (MI-006)	OIML R50-1/-2 : 2014(E)	Comments	Conclusions
<p>9.3. The instrument shall be accompanied by information on its operation, unless the simplicity of the measuring instrument makes this unnecessary. Information shall be easily understandable and shall include where relevant:</p> <ul style="list-style-type: none"> (a) rated operating conditions; (b) mechanical and electromagnetic environment classes; (c) the upper and lower temperature limit, whether condensation is possible or not, open or closed location; (d) instructions for installation, maintenance, repairs, permissible adjustments; (e) instructions for correct operation and any special conditions of use; (f) conditions for compatibility with interfaces, sub-assemblies or measuring instruments. 	<p>R50-1, 5.8 R50-1, 6.1.1</p>	<p>Consider the operating manual</p>	<p>Not covered</p>
<p>9.4. Groups of identical measuring instruments used in the same location or used for utility measurements do not necessarily require individual instruction manuals.</p>			<p>Not relevant</p>
<p>9.5. Unless specified otherwise in an instrument-specific annex, the scale interval for a measured value shall be in the form 1×10^n, 2×10^n, or 5×10^n, where n is any integer or zero. The unit of measurement or its symbol shall be shown close to the numerical value.</p>	<p>R50-1, 4.4.3.1</p>		<p>Covered</p>
<p>9.6. A material measure shall be marked with a nominal value or a scale, accompanied by the unit of measurement used.</p>			<p>Not relevant</p>
<p>9.7. The units of measurement used and their symbols shall be in accordance with the provisions of Union legislation on units of measurement and their symbols.</p>	<p>R50-1, 3.6</p>		<p>Covered</p>
<p>9.8. All marks and inscriptions required under any requirement shall be clear, non-erasable, unambiguous and non-transferable.</p>	<p>R50-1, 4.9, 4.9.4, 4.10</p>		<p>Covered</p>

Directive 2014/32/EU Essential requirements of Annex I and Annex VIII Automatic Weighing Instruments (MI-006)	OIML R50-1/-2 : 2014(E)	Comments	Conclusions
10. Indication of result			
10.1. Indication of the result shall be by means of a display or hard copy.	R50-1, 4.4		Covered
10.2. The indication of any result shall be clear and unambiguous and accompanied by such marks and inscriptions necessary to inform the user of the significance of the result. Easy reading of the presented result shall be permitted under normal conditions of use. Additional indications may be shown provided they cannot be confused with the metrologically controlled indications.	R50-1, 4.4.2		Covered
10.3. In the case of hard copy the print or record shall also be easily legible and non-erasable.	R50-1, 4.4.1, 4.4.7 R50-1, 5.7		Covered
10.4. A measuring instrument for direct sales trading transactions shall be designed to present the measurement result to both parties in the transaction when installed as intended. When critical in case of direct sales, any ticket provided to the consumer by an ancillary device not complying with the appropriate requirements of this Directive shall bear an appropriate restrictive information.			Not relevant
10.5. Whether or not a measuring instrument intended for utility measurement purposes can be remotely read it shall in any case be fitted with a metrologically controlled display accessible without tools to the consumer. The reading of this display is the measurement result that serves as the basis for the price to pay.			Not relevant

Directive 2014/32/EU Essential requirements of Annex I and Annex VIII Automatic Weighing Instruments (MI-006)	OIML R50-1/-2 : 2014(E)	Comments	Conclusions
11. Further processing of data to conclude the trading transaction			
11.1. A measuring instrument other than a utility measuring instrument shall record by a durable means the measurement result accompanied by information to identify the particular transaction, when: (a) the measurement is non-repeatable and; (b) the measuring instrument is normally intended for use in the absence of one of the trading parties.	R50-1, 4.4 R50-1, 5.5.4 R50-1, 5.7.1	There is no mandatory requirement to have a printer or a data storage device. <i>Instruments intended to be used in the absence of one trading party shall be equipped with a data storage device which records the measurement result accompanied by information to identify the particular transaction and reconstruct it later</i>	Partially covered
11.2. Additionally, a durable proof of the measurement result and the information to identify the transaction shall be available on request at the time the measurement is concluded.		There are no requirement concerning the information to identify a particular transaction	Not covered
12. Conformity evaluation A measuring instrument shall be designed so as to allow ready evaluation of its conformity with the appropriate requirements of this Directive.	R50-1, 6.2		Covered

Directive 2014/32/EU Essential requirements of Annex I and Annex VIII Automatic Weighing Instruments (MI-006)	OIML R50-1/-2 : 2014(E)	Comments	Conclusions
ANNEX MI-006 AUTOMATIC WEIGHING INSTRUMENTS			
The relevant essential requirements of Annex I, the specific requirements of this Annex and the conformity assessment procedures listed in Chapter I of this Annex, apply to automatic weighing instruments defined below, intended to determine the mass of a body by using the action of gravity on that body.	R50-1, 2.1.1	Slightly different wording	Covered

Directive 2014/32/EU Essential requirements of Annex I and Annex VIII Automatic Weighing Instruments (MI-006)	OIML R50-1/-2: 2014(E)	Comments	Conclusions
DEFINITIONS			
Automatic weighing instrument An instrument that determines the mass of a product without the intervention of an operator and follows a predetermined programme of automatic processes characteristic of the instrument.	R50-1, 2.1.2	Slightly different wording	Covered
Automatic catchweigher An automatic weighing instrument that determines the mass of pre-assembled discrete loads (for example prepackages) or single loads of loose material.			Not relevant
Automatic checkweigher An automatic catchweigher that subdivides articles of different mass into two or more subgroups according to the value of the difference of their mass and a nominal set-point.			Not relevant
Weight labeller An automatic catchweigher that labels individual articles with the weight value.			Not relevant
Weight/price labeller An automatic catchweigher that labels individual articles with the weight value, and price information.			Not relevant
Automatic gravimetric filling instrument An automatic weighing instrument that fills containers with a predetermined and virtually constant mass of product from bulk.			Not relevant
Discontinuous totaliser (totalising hopper weigher) An automatic weighing instrument that determines the mass of a bulk product by dividing it into discrete loads. The mass of each discrete load is determined in sequence and summed. Each discrete load is then delivered to bulk.			Not relevant

Directive 2014/32/EU Essential requirements of Annex I and Annex VIII Automatic Weighing Instruments (MI-006)	OIML R50-1/-2: 2014(E)	Comments	Conclusions
Continuous totaliser An automatic weighing instrument that continuously determines the mass of a bulk product on a conveyor belt, without systematic subdivision of the product and without interrupting the movement of the conveyor belt.	R50-1, 2.1.3	Slightly different title	Covered
Rail-weighbridge An automatic weighing instrument having a load receptor inclusive of rails for conveying railway vehicles.			Not relevant

Directive 2014/32/EU Essential requirements of Annex I and Annex VIII Automatic Weighing Instruments (MI-006)	OIML R50-1 2014(E)	Comments	Conclusions
SPECIFIC REQUIREMENTS			
Chapter I – Requirements common to all types of automatic weighing instruments			
1. <i>Rated Operating Conditions</i> The manufacturer shall specify the rated operating conditions for the instrument as follows:	R50-1, 4.2 R50-1, 6.1.1		Covered
1.1. For the measurand: The measuring range in terms of its maximum and minimum capacity.	R50-1, 2.3.4 R50-1, 6.1.1		Covered
1.2. For the electrical supply influence quantities: In case of AC voltage supply: the nominal AC voltage supply, or the AC voltage limits. In case of DC voltage supply: the nominal and minimum DC voltage supply, or the DC voltage limits.	R50-1, 3.7.4.3 R50-1, 3.7.4.3 R50-1, 6.1.1		Covered
1.3. For the mechanical and climatic influence quantities: The minimum temperature range is 30°C unless specified otherwise in the following chapters of this Annex. The mechanical environment classes according to Annex I, paragraph 1.3.2 are not applicable. For instruments which are used under special mechanical strain, e.g. instruments incorporated into vehicles, the manufacturer shall define the mechanical conditions of use.	R50-1, 3.7.4.1 R50-1, 6.1.1		Covered
1.4. For other influence quantities (if applicable): The rate(s) of operation. The characteristics of the product(s) to be weighed	R50-1, 3.5 R50-1, 6.1.1		Covered

Directive 2014/32/EU Essential requirements of Annex I and Annex VIII Automatic Weighing Instruments (MI-006)	OIML R50-1 2014(E)	Comments	Conclusions
<p>2. <i>Permissible effect of disturbances – Electromagnetic environment</i></p> <p>The required performance and the critical change value are given in the relevant Chapter of this Annex for each type of instrument.</p>		See chapter V of annex VIII	
3. <i>Suitability</i>			
3.1. Means shall be provided to limit the effects of tilt, loading and rate of operation such that maximum permissible errors (MPEs) are not exceeded in normal operation.	R50-1, 4.1, 4.7 & 4.8 R50-1, 4.2		Covered
3.2. Adequate material handling facilities shall be provided to enable the instrument to respect the MPEs during normal operation.	R50-1, 4.8.1		Covered
3.3. Any operator control interface shall be clear and effective.	R50-1, 4.3.4 R50-1, 4.4.1		Covered
3.4. The integrity of the display (where present) shall be verifiable by the operator.	R50-1, 5.4		Covered
3.5. Adequate zero setting capability shall be provided to enable the instrument to respect the MPEs during normal operation.	R50-1, 3.7.3, 3.7.5.4 & 4.5 R 50-2, 5.4.3 & 5.4.4 R50-1, 3.8.2 & R50-2, 9.1 R50-1, 3.8.3 & R50-2, 9.1.1 R50-1, 3.8.4 & R50-2, 9.1.2		Covered
3.6. Any result outside the measurement range shall be identified as such, where a printout is possible.	R50-1, 4.3.6		Covered

Directive 2014/32/EU Essential requirements of Annex I and Annex VIII Automatic Weighing Instruments (MI-006)	OIML R50-1 2014(E)	Comments	Conclusions
<p>4. Conformity assessment</p> <p>The conformity assessment procedures referred to in Article 17 that the manufacturer can choose between are:</p> <p>For mechanical systems: B+D or B+E or B+F or D1 or F1 or G or H1.</p> <p>For electromechanical instruments: B+D or B+E or B+F or G or H1.</p> <p>For electronic systems or systems containing software: B+D or B+F or G or H1.</p>			Not covered
Chapter V – Continuous Totalisers			
<p>1. <i>Accuracy Classes</i></p> <p>Instruments are divided into three accuracy classes as follows: 0.5, 1, 2.</p>	R50-1, 3.1		Covered
<p>2. <i>Measurement Range</i></p>			
<p>2.1. The manufacturer shall specify the measurement range, the ratio between the minimum net load on the weighing unit and the maximum capacity, and the minimum totalised load.</p>	R50-1, 3.4 R50-1, 3.5 R50-1, 6.1.1	No measuring range, but definition by Max and flow rate values	Covered
<p>2.2. The minimum totalised load Σ_{\min} shall not be less than</p> <p>800 d for class 0.5, 400 d for class 1, 200 d for class 2.</p> <p>Where d is the totalisation scale interval of the general totalisation device.</p>	R50-1, 3.4		Covered

Directive 2014/32/EU Essential requirements of Annex I and Annex VIII Automatic Weighing Instruments (MI-006)		OIML R50-1 2014(E)	Comments	Conclusions	
3 <i>MPE</i> <div>Table 8</div>		R50-1, 3.2.1 - Table 1	Covered provided that the MPE is not rounded. The sentence in 3.2.1 of R 50 is not in annex VIII of MID.	Partially covered	
	Accuracy class				MPE for totalised load
	0.5				± 0,25 %
	1				± 0,5 %
	2				± 1,0 %
4 <i>Speed of the belt</i> The speed of the belt shall be specified by the manufacturer. For single-speed beltweighers, and variable-speed beltweighers having a manual speed setting control, the speed shall not vary by more than 5 % of the nominal value. The product shall not have a different speed than the speed of the belt.		R50-1, 6.1.1 R50-1, 4.8.1.3 R50-1, 4.8.1 d)		Covered	
5 <i>General Totalisation Device</i> It shall not be possible to reset the general totalisation device to zero.		R50-1, 4.3.2		Covered	
6. <i>Performance under influence factors and electromagnetic disturbances</i>					
6.1. The MPE due to influence factor, for a load not less than the Σ_{min} , shall be 0,7 times the appropriate value specified in Table 8, rounded to the nearest totalisation scale interval (d).		R50-1, 3.2.2		Covered	
6.2. The critical change value due to a disturbance shall be 0.7 times the appropriate value specified in Table 8, for a load equal to Σ_{min} , for the designated class of the beltweigher; rounded up to the next higher totalisation scale interval (d).		R50-1, 2.4.5.4	Term “significant fault” is used	Covered	